

CORRIEGENDUM

Tender no-CPC/EQP-RD/13-14/15 (Second Call) FOR Supply of Radio Diagnosis Equipment

S.N o.	ITEM CODE		EXISTING	MODIFICATION
1	RD-8 (50 KW high frequency x ray machine with IITV)	Page no. 79 line no. 1	<u>SPECIFICATIONS OF 80KW HIGH FREQUENCY X-RAY MACHINE WITH I I TV</u>	<u>Modified as- SPECIFICATIONS OF 50KW HIGH FREQUENCY X-RAY MACHINE WITH I I TV</u>
		Page no. 79 line no. 3	<u>X-Ray Generator:</u> - A 800mA or more, 150 KV, 80 KW High Frequency X-Ray Generators for General Radiography & Fluoroscopy.	<u>Modified as- "X-Ray Generator:</u> - A 800mA or more, 125 KV, 50 KW High Frequency X-Ray Generators for General Radiography & Fluoroscopy.
		Page no. 81 line no. 6	<u>IITV System and Digital Fluoro Memory: (Quote as optional)</u>	<u>Word "(Quote as optional)" deleted.</u>

Jah

Lu
10-07-13

Pa

S
10-7-13

Jes
10-07-2013